

Certificate of Analysis

Total Organic Carbon Standard

TOC Standard 0.1mg/L C as KHP

Product No:	RTOCK06	Date of Measurement:	19/04/2023
Lot No:	TOCK0623D1	Date of Sample Receipt:	19/04/2023
Expiry Date*:	02/06/2023	Date of Manufacture*:	19/04/2023

Specification:	Mean Measured Value:
0.095 - 0.105 mg/L C as KHP	0.098 mg/L C as KHP

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Tested in accordance with In-House Test Method. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.


Storage Conditions:

Product must be stored at 2 to 8°C when not in use.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :
21/04/2023

Senior QC Technician
MALONE Eileen

A handwritten signature in black ink that reads "Eileen Malone".

All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL1