

Shannon Free Zone, Shannon, Co. Clare, Ireland Tel: +353 61 472622 Fax: +353 61 472642 Email:sales@reagecon.ie

www.reagecon.com

CERTIFICATE OF GRAVIMETRIC PREPARATION

PRODUCT: IC Multi-Element Standard (3 elements)

PRODUCT No.: ICA3-MIX3-A

MATRIX: H_2O

LOT NO.: ICA3M3B24K1

DATE OF PREPARATION: 22nd October 2024

EXPIRY DATE: 28th October 2026

DENSITY VALUE: 0.998 g/ml @ 20°C

PREPARATION OF STANDARD:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Elements	Nominal mg/kg	Actual mg/kg	Actual mg/L @ 20°C
Chloride (as Cl ⁻)	10.02	10.01	9.99
Fluoride (as F ⁻)	10.02	10.01	9.99
Sulphate (as SO ₄ ² -)	10.02	10.01	9.99

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as \pm 0.2 %. All values are verified by IC analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

TRACEABILITY IN THE PRODUCTION OF THIS STANDARD:

This product was prepared gravimetrically on a mass/mass basis, using balances calibrated by Reagecon engineers with mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity and linearity.

TEST METHOD:

The mean result of this standard was verified using a calibrated IC system according to an inhouse test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: QC Supervisor

Date: 16th October 2024

Atton

This certificate must not be reproduced except in full.