

# Certificate of Conformity

## Water and Synthetic Fresh Water Standards

### Purified Water

Product No:  
Lot No:  
Expiry date :

H2O5  
H2O25C2  
28/03/2026

Date of measurement:

31/03/2025

#### Physical Description:

Clear, Colourless, Odourless Liquid

Formula: H<sub>2</sub>O Molecular Weight: 18.02

CAS #: 7732-18-5

#### Specification:

Conforms to USP 36 for Packaged Purified Water

Max. specific conductance @25°C 1µS/cm-1(when first tested

pH: 5.0 - 7.0

Ammonium: Pass

Calcium: Pass

Carbon Dioxide: Pass

Chloride: Pass

Sulphate: Pass

Oxidisable Substances: Pass

Fluoride: <0.5ppm

TOC <500ppb

As per chapter 1231, this Purified Water meets the requirements of "bulk packaging for commercial use elsewhere". This product has not been sterilised. It is the users responsibility to assure fitness for use of this packaged article when used in manufacturing, clinical, or analytical applications where the pure bulk form of the water is indicated.

This is to certify that this product conforms to the specification given above when used as directed.

Date of issue of the certificate:

01/04/2025

*gabriel*

QC Technician **Gabriel Ondiano**

Complementary information relative to this product is available at [www.reagecon.com](http://www.reagecon.com)  
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-25A1

Tested by Reagecon Quality Control  
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.  
Shannon Free Zone, Shannon, Co. Clare, Ireland.  
Tel +353 61 472622, Fax: +353 61 472642  
Email: [sales@reagecon.ie](mailto:sales@reagecon.ie), [www.reagecon.com](http://www.reagecon.com)