



CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global Certificate Number. QMS42099

Δ/0627 Catalogue Number Lot Number 2539531

Description Acetonitrile HPLC HPLC for Far UV/gradient analysis

CAS Number 75-05-8 Quality Test/Release Date 29/May/2025

Expiry Phrase Use within 3 years of opening

China Country of Origin **Declaration of Origin** Synthetic

Not directly derived from or manufactured with any animal byproducts in any way BSE/TSE

(Including but not limited to fermentation or nutrient broth, catalysts, enzymes).

Result Name	Units	Specifications	Test Value	
Absorbance @ 195 nm	A.U.	<= 0.1	0.1	
Absorbance @ 200 nm	A.U.	<= 0.05	0.042	
Absorbance @ 210 nm	A.U.	<= 0.02	0.018	
Absorbance @ 220 nm	A.U.	<= 0.01	0.009	
Absorbance @ 230 nm	A.U.	<= 0.005	0.004	
Absorbance @ 240 nm	A.U.	<= 0.005	0.002	
Absorbance @ 250 nm	A.U.	<= 0.005	0.001	
Acidity	meq/g	<= 0.0008	0.00002	
Assay (GC)	%	>= 99.9	100.00	
Chemical form		Mobile liquid	Mobile liquid	
Gradient analysis (max. eluted peak)	A.U.	<= 0.01	None Detected	
Residue after evaporation (ppm)	ppm	<= 2	None Detected	
Visual colour		Clear colourless	Clear colourless	
Water	%	<= 0.01	0.003	

Additional Information Meets the current analytical specification of Ph.Eur. Filtered to 0.2 micron

Ashok Ganatra

A.Ganatra

Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.