

## CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global  
Certificate Number. QMS42099

Catalogue Number	41004
Lot Number	2173376
Description	Diethyl ether HPLC HPLC
CAS Number	60-29-7
Quality Test/Release Date	01/Mar/2021
Expiry Phrase	Use within 2 yrs of opening
Country of Origin	India

Result Name	Units	Specifications	Test Value
Absorbance @ 208 nm	A.U.	<= 1	0.133
Absorbance @ 210 nm	A.U.	<= 0.523	0.105
Absorbance @ 230 nm	A.U.	<= 0.301	0.044
Absorbance @ 250 nm	A.U.	<= 0.097	0.012
Absorbance @ 270 nm	A.U.	<= 0.046	0.002
Absorbance @ 280 nm	A.U.	<= 0.022	0.001
Absorbance @ 400 nm	A.U.	<= 0.009	0
Acidity	meq/g	<= 0.0002	0.00005
Alkalinity (meq/g)	meq/g	<= 0.0002	None Detected
Assay (GC)	%	>= 98	99.98
Chemical form		Mobile liquid	Mobile liquid
Colour	APHA	<= 10	<5
Non volatile matter	%	<= 0.0003	0.0001
Peroxide (ppm)	ppm	<= 1	<0.15
Visual colour		Clear colourless	Clear colourless
Water	%	<= 0.05	0.01

**Additional Information** 2% Ethanol added as a stabilizer



Ashok Ganatra  
Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.