

Fisher Scientific UK Ltd. Bishop Meadow Road Loughborough, LE11 5RG, United Kingdom

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global Certificate Number. QMS42099

Catalogue Number	T/3606		
Lot Number	2535091		
Description	2,2,4-Trimethylpentane, (iso-octane) HPLC HPLC 540-84-1		
CAS Number			
Quality Test/Release Date	28/Apr/2025		
Expiry Phrase	Use within 5 yrs of opening		
Country of Origin	Germany		
 Result Name	Units	Specifications	Test Value
Absorbance @ 200 nm	A.U.	<= 0.4	0.162
Absorbance @ 210 nm	A.U.	<= 0.2	0.054
Absorbance @ 220 nm	A.U.	<= 0.1	0.017
Absorbance @ 230 nm	A.U.	<= 0.02	0.007
Absorbance @ 240 nm	A.U.	<= 0.005	0
Absorbance @ 250 nm	A.U.	<= 0.005	0
Acidity/alkalinity (meq/g)	meq/g	<= 0.0002	0.00004
Assay (GC)	%	>= 99.5	99.90
Chemical form		Mobile liquid	Mobile liquid
Residue after evaporation (ppm)	ppm	<= 2	None Detected
Visual colour		Clear colourless	Clear colourless
Water	%	<= 0.01	0.003

Additional Information Filtered to 0.2 micron

A.Ganatra

Ashok Ganatra Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.