

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global
Certificate Number. QMS42099

Catalogue Number	022904
Lot Number	2537763
Description	Tetrahydrofuran, UV, HPLC Grade, 99.7+% unstab. HPLC HPLC
CAS Number	109-99-9
Quality Test/Release Date	12/May/2025
Expiry Phrase	Use within 1 year of packing date
Country of Origin	Austria

Result Name	Units	Specifications	Test Value
Absorbance @ 212 nm	A.U.	<= 1	0.892
Absorbance @ 225 nm	A.U.	<= 0.50	0.288
Absorbance @ 250 nm	A.U.	<= 0.17	0.097
Absorbance @ 350 nm	A.U.	<= 0.01	0
Assay (GC)	%	>= 99.7	100.00
Chemical form		Mobile liquid	Mobile liquid
Peroxide Content (at time of packaging)	%	<= 0.015	0.0029
Residue after evaporation (ppm)	ppm	<= 5	None Detected
Visual colour		Clear colourless	Clear colourless
Water	%	<= 0.03	0.02

Additional Information Filtered to 0.2 micron



Ashok Ganatra
Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.