

TRACEABLE TO NIST STD

Fisher Scientific UK Ltd. 755 US Highway Bridgewater, NJ 08807 202, United States of America

SRM 84I

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by Intertek Global Certificate Number. CERT-0120633

Catalogue Number	SA56		
Lot Number	241644		
Description	HYDROCHLORIC ACID, 6N		
CAS Number	7732-18-5		
Quality Test/Release Date	09/Aug/2024		
Suggested retest date	28/Feb/2026		
Country of Origin	United States		
Declaration of Origin	Inorganic-non animal		
	No animal products are used as starting raw material ingredients, or used in		
BSE/TSE	processing, including lubricants, processing aids, or any other material that might		
	migrate to the finished product.		
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless solution
COLOR	APHA	<= 10	<10
NORMALITY		Inclusive Between 5.95 - 6.05	6.025

qui

NIST SRM = SRM 84I;SRM 732e

Matthew Micek QC Supervisior

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.