

Fisher Scientific UK Ltd. 1 Reagent Lane Fair Lawn, NJ 07410, United States of America

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by Intertek Global Certificate Number. CERT-0120633

Catalogue Number	SA54			
Lot Number	251267			
Description	HYDROCHLORIC ACID, 0.1N, CERTIFIED			
CAS Number	7647-01-0			
Quality Test/Release Date	15/May/2025			
Suggested retest date	15/May/2027			
Country of Origin	United States			
Declaration of Origin	Inorganic-non animal			
BSE/TSE	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.			
Result Name	Units	Specifications	Test Value	
APPEARANCE		REPORT	Clear, colorless liquid	
COLOR	APHA	<= 10	<10	
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST	
NORMALITY	NORMAL	Inclusive Between 0.0995 - 0.1005 0.0999		
Stoichiometry NIST SRMs	NIST SRM	IN SRM 84I;SRM 723e	SRM 84I	

Additional Information This material is free from genetically modified substances.

Cut Salyn

Harout Sahagian QC Manager

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.