

## CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by Intertek Global Certificate Number. CERT-0120633

|                           |   |
|---------------------------|---|
| Catalogue Number          | E196  |
| Lot Number                | 261026  |
| Description               | ETHYL ACETATE - OPTIMA  |
| CAS Number                | 141-78-6  |
| Quality Test/Release Date | 10/Jun/2026   |
| Suggested retest date     | 09/Jun/2031   |
| Country of Origin         | United States   |
| Declaration of Origin     | Organic - non animal  |
| BSE/TSE                   | No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product. |

| Result Name                  | Units      | Specifications                    | Test Value                                       |
|------------------------------|------------|-----------------------------------|--|
| APPEARANCE                   |            | REPORT                            | Clear, colorless liquid free of suspended matter |
| ASSAY                        | %          | >= 99.9                           | 99.98  |
| COLOR                        | APHA       | <= 5                              | 5  |
| DENSITY AT 25 DEGREES C      | GM/ML      | Inclusive Between 0.893 - 0.895   | 0.894  |
| EVAPORATION RESIDUE          | ppm        | <= 1                              | <0.1   |
| FLUORESCENCE BACKGROUND      | ppb        | <= 1                              | <1   |
| IDENTIFICATION               | PASS/FAIL  | = PASS TEST                       | PASS TEST  |
| OPTICAL ABS AT 255 NM        | ABS. UNITS | <= 1.00                           | 0.53   |
| OPTICAL ABS AT 260 NM        | ABS. UNITS | <= 0.12                           | 0.09   |
| OPTICAL ABS AT 270 NM        | ABS. UNITS | <= 0.015                          | 0.002  |
| PESTICIDE RESIDUE ANALYSIS   | NG/L       | <= 10                             | <1   |
| REFRACTIVE INDEX @ 25 DEG C  |            | Inclusive Between 1.3680 - 1.3710 | 1.3704   |
| SUBSTANCES DARKENED BY H2SO4 | PASS/FAIL  | = PASS TEST                       | PASS TEST  |
| TITRATABLE ACID              | MEQ/G      | <= 0.0009                         | <0.00001   |
| WATER (H2O)                  | %          | <= 0.02                           | 0.002  |



Ash Frey  
Supervisor, Quality Control

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.