

Fisher Scientific UK Ltd. Bishop Meadow Road Loughborough, LE11 5RG, United Kingdom

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global Certificate Number. QMS42099

Catalogue Number	A/0627		
Lot Number	2543706		
Description	Acetonitrile HPLC HPLC for Far UV/gradient analysis		
CAS Number	75-05-8		
Quality Test/Release Date	04/Jul/2025		
Expiry Phrase	Use within 3 years of opening		
Country of Origin	China		
Declaration of Origin	Synthetic		
BSE/TSE	Not directly derived from or manufactured with any animal byproducts in any way (Including but not limited to fermentation or nutrient broth, catalysts, enzymes).		
	(Including	but not inflited to termentation	n or nument broth, catalysts, enzymes).
Result Name	Units	Specifications	Test Value
Absorbance @ 195 nm	A.U.	<= 0.1	0.1
Absorbance @ 200 nm	A.U.	<= 0.05	0.041
Absorbance @ 210 nm	A.U.	<= 0.02	0.019
Absorbance @ 220 nm	A.U.	<= 0.01	0.01
Absorbance @ 230 nm	A.U.	<= 0.005	0.005
Absorbance @ 240 nm	A.U.	<= 0.005	0.003
Absorbance @ 250 nm	A.U.	<= 0.005	0.002
Acidity	meq/g	<= 0.0008	0.00001
Assay (GC)	%	>= 99.9	100.00
Chemical form		Mobile liquid	Mobile liquid
Gradient analysis (max. eluted peak)	A.U.	<= 0.01	None Detected
Residue after evaporation (ppm)	ppm	<= 2	None Detected
Visual colour		Clear colourless	Clear colourless
Water	%	<= 0.01	0.004

Additional Information Meets the current analytical specification of Ph.Eur. Filtered to 0.2 micron

A.Ganatra

Ashok Ganatra Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.