

Fisher Scientific UK Ltd. Bishop Meadow Road Loughborough, LE11 5RG, United Kingdom

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global Certificate Number. QMS42099

Catalogue Number	A/0630			
Lot Number	2534663			
Description	Acetonitrile Fluorescence HPLC For fluorescence detection			
CAS Number	75-05-8			
Quality Test/Release Date	14/Apr/2025			
Expiry Phrase	Use within 3 years of opening			
Country of Origin	China			
Result Name	Units	Specifications	Test Value	
Absorbance @ 200 nm	A.U.	<= 0.1	0.039	
Absorbance @ 205 nm	A.U.	<= 0.05	0.026	
Absorbance @ 210 nm	A.U.	<= 0.035	0.017	
Absorbance @ 220 nm	A.U.	<= 0.025	0.008	
Absorbance @ 230 nm	A.U.	<= 0.01	0.004	
Absorbance @ 240 nm	A.U.	<= 0.005	0.002	
Absorbance @ 250 nm	A.U.	<= 0.005	0.001	
Acidity	meq/g	<= 0.0008	0.00001	
Assay (GC)	%	>= 99.9	100.00	
Chemical form		Mobile liquid	Mobile liquid	
Fluorescence		Must be 'Pass'	Pass	
Residue after evaporation (ppm)	ppm	<= 2	None Detected	
Visual colour		Clear colourless	Clear colourless	
Water	%	<= 0.02	0.002	

A.Ganatra

Ashok Ganatra Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.