

Thermo Fisher Scientific Ward Hill, MA-Sites

 30 Bond Street
 220 Neck Road
 38 Shelley Road

 Ward Hill
 Ward Hill
 Ward Hill

 MA 01835
 MA 01835
 MA 01835

## **CERTIFICATE OF ANALYSIS**

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by TUV SUD America Inc.

Certificate Number. 951001163

Catalogue Number	036645
Lot Number	S23L022

Description Water, Reagent (Deionized water), ACS

CAS Number 7732-18-5
Quality Test/Release Date 23/Jun/2025
Suggested retest date 23/Jun/2027
Country of Origin United States

Result Name	Units	Specifications	Test Value
Chloride content	ppm	0.4ppm max	0.0001
Conductivity	Siemens/cm	0.000002 Siemens per cm @ $25^{\circ}$ C max	0.00000056
Impurity content	ppm	(Nitrate, NO₃) : 0.4ppm max	<0.00002
Impurity content	ppm	(Phosphate, PO₄) : 1ppm max	<0.00002
Impurity content	ppm	(Sulphate, SO <sub>4</sub> ) : 1ppm	<0.00005
Impurity content	ppm	Silicate (as SiO <sub>2</sub> ): 0.01ppm max	<0.001
Impurity content		Substances reducing permanganate : Passes test	Passes test
Trace Metals	ppm	Heavy metals (as Pb) : 0.01ppm max	<0.01

John Astarita

John Astarita
Director, Quality

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.