

Catalog Number	13-678-12D	Mfg. Date	05/11/2026
Lot Number	13126044	Sample Id	N/A
Product Description	Fisher 5 mL pipette, bulk pack		

### Component Materials

Pipette	Polystyrene, meets USP Cytotoxicity Test <87>.
Ink	Blue / black, meets USP Cytotoxicity Test <87>.
Mouthpiece Plug	Polyester, meets USP Cytotoxicity Test <87>.

### Sterilization:

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11 137 Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization. Sterility Assurance Level: SAL  $10^{-6}$

### Pyrogens:

Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing". The acceptance level for product is  $\leq 0.10$  EU/ml or  $\leq 4$  EU/device.

### DNase/RNase:

Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is  $10^{-7}$  Kunitz units/uL for DNase and  $10^{-9}$  Kunitz units/uL for RNase

### BSE/TSE:

This product is deemed animal free by virtue of not containing materials of animal origin and/or complies with the latest revision of EMA/410/01 section 6.4.

### Volumetric Accuracy:

Serological pipets are accurate to +/- 2% at full volume in compliance with ASTM E934, "Standard Specification for Serological Pipet, Disposable Plastic" and ISO 12771, "Plastics laboratory ware - Disposable serological pipettes".

### Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing, Next 2 digits -Year of manufacture; Last 3 digits - Run number for that year

### Expiration Statement:

This product has an expiration date of 5 years from the date of manufacture for the attributes contained within this certificate.

Quality Assurance Department  
 Fisher Scientific Company L.L.C.

This document has been produced electronically and is valid without a signature.