

Fisher Scientific UK Ltd. Bishop Meadow Road Loughborough, LE11 5RG, United Kingdom

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global Certificate Number. QMS42099

Catalogue Number	26828			
Lot Number	2543908			
Description	Methanol			
CAS Number	67-56-1			
Quality Test/Release Date	04/Jul/2025			
Expiry Phrase	Use within 3 years of opening			
Country of Origin	Trinidad and Tobago			
Result Name	Units	Specifications	Test Value	
Absorbance @ 210 nm	A.U.	<= 0.2	0.095	
Absorbance @ 220 nm	A.U.	<= 0.1	0.015	
Absorbance @ 230 nm	A.U.	<= 0.05	0.005	
Absorbance @ 240 nm	A.U.	<= 0.02	0.003	
Absorbance @ 250 nm	A.U.	<= 0.005	0.002	
Absorbance @ 260 nm	A.U.	<= 0.005	0.003	
Acidity/alkalinity (meq/g)	meq/g	<= 0.0002	0.00003	
Assay (GC)	%	>= 99.8	100.00	
Chemical form		Mobile liquid	Mobile liquid	
Residue after evaporation (ppm)	ppm	<= 5	None Detected	
Visual colour		Clear colourless	Clear colourless	
Water	%	<= 0.05	0.02	
Wavelength @ 1 A.U.	nm	<= 205	201.8	

A.Ganatra

Ashok Ganatra Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.