

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by Intertek. Certificate Number. 2317548

Catalogue Number	29559
Lot Number	A0475098
Description	Azur eosin methylene-blue,pure,solution according to Giemsa
CAS Number	51811-82-6
Quality Test/Release Date	11/Aug/2025
Suggested retest date	11/Aug/2030
Country of Origin	UNITED KINGDOM
Declaration of Origin	synthetic
BSE/TSE	Not directly derived from or manufactured with any animal byproducts in any way (Including but not limited to fermentation or nutrient broth, catalysts, enzymes).

Result Name	Units	Specifications	Test Value
Appearance (Color)		Dark blue to dark purple	Dark blue
Appearance (Form)		Solution	Solution
UV Lambda max	nm	523 to 526 (c=0.1, MeOH) Eosine	523 (c=0.1, MeOH) Eosine
UV Lambda max	nm	646 to 651 (c=0.1, MeOH) Methylene blue	647 (c=0.1, MeOH) Methylene blue
Additional info		Giemsa stain powder in MeOH, Glycerol mixture	Giemsa stain powder in MeOH, Glycerol mixture



Geert Torfs
Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.