

Catalog Number	FB1495916	Mfg. Date	09/02/2025
Lot Number	24425072	Sample Id	N/A
Product Description	Fisher 5ml PP round bottom test tube w/o cap, bagged, non-sterile		

### Component Materials

Tube	Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures.
Cap	Virgin High Density Polyethylene, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG req.) color concentrate.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure	Result
Leak Test	Pass
Centrifugation Test	Pass
Visual Attributes	Pass
Packaging	Pass

#### Sterilization:

This lot has been irradiated and dosimetrically released based on the Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization. Sterility Assurance Level: SAL  $10^{-5}$

#### Pyrogens:

Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing". The acceptance level for product is  $\leq 0.10$  EU/ml or  $\leq 4$  EU/device.

#### DNase/RNase:

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is  $10^{-7}$  Kunitz units/uL for DNase and  $10^{-9}$  Kunitz units/uL for RNase.

#### Expiration Statement:

"This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate."

#### Lot Number Designation:

8 Digit Lot Number: First digit - Last number of year of manufacture, Next 3 digits - Julian date, start of manufacturing, Last 3 digits - Batch identification

Quality Assurance Department  
 Fisher Scientific Company L.L.C.

This document has been produced electronically and is valid without a signature.

This certificate has been compiled with data that has been provided by the manufacturer of the product.  
 Fisher Scientific Channel has not independently verified such data.

This certificate has been compiled with data that has been provided by the manufacturer of the product.  
Fisher Scientific Channel has not independently verified such data.