

## CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by Intertek. Certificate Number. 2317548

Catalogue Number	40691
Lot Number	A0471262
Description	Dichloromethane, 99.9%, for biochemistry, stabilized with approx. 50 ppm amylene, AcroSeal®
CAS Number	75-09-2
Quality Test/Release Date	01/Apr/2025
Suggested retest date	01/Apr/2028
Country of Origin	UNITED KINGDOM
Declaration of Origin	synthetic
BSE/TSE	Not directly derived from or manufactured with any animal byproducts in any way (Including but not limited to fermentation or nutrient broth, catalysts, enzymes).

Result Name	Units	Specifications	Test Value
Appearance (Form)		Clear liquid	Clear liquid
Color scale	APHA	≤10	10
Infrared spectrum		Conforms	Conforms
GC	%	≥99.85	99.9
Water	%	≤0.02	0.0024
Acidity	meq/g	≤0.0002	≤0.0002
Alkalinity	meq/g	≤0.0002	≤0.0002
Residue after evaporation	%	≤0.001	≤0.001
UV		at 400 nm A: ≤0.01	at 400 nm A: 0.0019
UV		at 260 nm A: ≤0.02	at 260 nm A: 0.0052
UV		at 250 nm A: ≤0.05	at 250 nm A: 0.0135
UV		at 240 nm A: ≤0.3	at 240 nm A: 0.0925
UV		at 235 nm A: ≤1.0	at 235 nm A: 0.3013
Stabilizer		Approx. 50 ppm Amylene	Approx. 50 ppm Amylene



Geert Torfs  
Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.