



## **CERTIFICATE OF ANALYSIS**

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global Certificate Number. QMS42099

M/4058 Catalogue Number Lot Number 2542222

Description Methanol HPLC HPLC for gradient analysis

CAS Number 67-56-1 Quality Test/Release Date 20/Jun/2025

Expiry Phrase Use within 3 yrs of opening Country of Origin Trinidad and Tobago

**Declaration of Origin** Synthetic

Not directly derived from or manufactured with any animal byproducts in any way BSE/TSE

(Including but not limited to fermentation or nutrient broth, catalysts, enzymes).

Result Name	Units	Specifications	Test Value	
Absorbance @ 210 nm	A.U.	<= 0.2	0.151	
Absorbance @ 220 nm	A.U.	<= 0.1	0.035	
Absorbance @ 230 nm	A.U.	<= 0.05	0.016	
Absorbance @ 240 nm	A.U.	<= 0.01	0.009	
Absorbance @ 250 nm	A.U.	<= 0.005	0.005	
Absorbance @ 260 nm	A.U.	<= 0.005	0.005	
Acidity	meq/g	<= 0.0002	0.00003	
Alkalinity (meq/g)	meq/g	<= 0.0002	None Detected	
Assay (GC)	%	>= 99.9	100.00	
Chemical form		Mobile liquid	Mobile liquid	
Gradient analysis (max. eluted peak)	A.U.	<= 0.001	None Detected	
Residue after evaporation (ppm)	ppm	<= 5	0.76	
Visual colour		Clear colourless	Clear colourless	
Water	%	<= 0.05	0.031	

Additional Information Filtered to 0.2 micron

Ashok Ganatra

Supervisor, QC

A. Ganatra

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.