

## CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by LRQA. Certificate Number. 10551915

Catalogue Number	038993
Lot Number	A0458176
Description	Carbon disulfide, HPLC Grade, 99.8+%
CAS Number	75-15-0
Quality Test/Release Date	30/May/2025
Suggested retest date	30/May/2028
Country of Origin	Great Britain
Declaration of Origin	Synthetic
BSE/TSE	Not directly derived from or manufactured with any animal byproducts in any way (including but not limited to fermentation or nutrient broth, catalysts, enzymes).

Result Name	Units	Specifications	Test Value
Color (APHA)		≤10	<10
Assay (GC)	%	≥99.8%	>=99.98
Appearance (Color)		Clear colorless	Clear colourless
Form		Liquid	Liquid
Residual water	%	≤0.10%	0.01
Optical Absorbance at		380nm: ≤1.3 AU (non-U.S. specification)	1.029
Optical Absorbance at		390nm: ≤0.5 AU (non-U.S. specification)	0.266
Optical Absorbance at		400nm: ≤0.10 AU (non-U.S. specification)	0.067
Optical Absorbance at		410nm: ≤0.05 AU	0.019
Optical Absorbance at		420nm: ≤0.01 AU (non-U.S. specification)	0.008
Residue on Evaporation	%	≤0.0005% (≤5ppm)	<=0.0001



Paul Coleman  
Site Quality Manager

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.