

## CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by INTERTEK SAI Global  
Certificate Number. QMS42099

Catalogue Number	D/2506
Lot Number	2535405
Description	Diethyl ether HPLC HPLC
CAS Number	60-29-7
Quality Test/Release Date	22/Apr/2025
Expiry Phrase	Use within 2 yrs of opening
Country of Origin	Germany

Result Name	Units	Specifications	Test Value
Absorbance @ 210 nm	A.U.	<= 0.3	0.052
Absorbance @ 220 nm	A.U.	<= 0.2	0.033
Absorbance @ 230 nm	A.U.	<= 0.125	0.019
Absorbance @ 240 nm	A.U.	<= 0.075	0.008
Absorbance @ 250 nm	A.U.	<= 0.05	0.004
Absorbance @ 260 nm	A.U.	<= 0.025	0.002
Absorbance @ 270 nm	A.U.	<= 0.01	0.002
Absorbance @ 280 nm	A.U.	<= 0.005	0.001
Acidity/alkalinity (meq/g)	meq/g	<= 0.0002	0.00003
Assay (GC)	%	>= 99	99.98
Chemical form		Mobile liquid	Mobile liquid
Residue after evaporation (ppm)	ppm	<= 2	None Detected
Visual colour		Clear colourless	Clear colourless
Water	%	<= 0.02	0.004
Wavelength @ 1 A.U.	nm	<= 205	201.6

**Additional Information** Filtered to 0.2 micron. Stabilised with Ethanol.



Ashok Ganatra  
Supervisor, QC

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research use or further manufacturing. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above.