

Thermo Fisher Scientific Ward Hill, MA-Sites

 30 Bond Street
 220 Neck Road
 38 Shelley Road

 Ward Hill
 Ward Hill
 Ward Hill

 MA 01835
 MA 01835
 MA 01835

CERTIFICATE OF ANALYSIS

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by TUV SUD America Inc.

Certificate Number. 951001163

Catalogue Number 038705 Lot Number Q14L035

Description Phenolphthalein, ACS

CAS Number 77-09-8
Quality Test/Release Date 14/Apr/2025
Suggested retest date 14/Apr/2030
Country of Origin India
Declaration of Origin Synthetic

 Result Name
 Units
 Specifications
 Test Value

 Comment
 Visual transition interval from pH 8.0 (colorless) to 10.0 (red)
 Conforms

 Comment
 Clarity of alcohol solution P.T.
 Passes test

John Astarita

John Astarita Director, Quality

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third-party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.