

| | | | |
|----------------------------|---|------------------|------------|
| Catalog Number | 0644321 | Mfg. Date | 02/25/2026 |
| Lot Number | 05026010 | Sample Id | N/A |
| Product Description | Fisher 50 mL polypropylene centrifuge tube, plug seal cap, bulk | | |

Component Materials

| | |
|------|--|
| Tube | Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures. |
| Cap | Virgin High Density Polyethylene, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG req.) color concentrate. |

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

| Test/Procedure | Result |
|--------------------------------|---------------|
| Leak Test | Pass |
| Centrifugation Test (9400 x g) | Pass |
| Visual Attributes | Pass |
| Packaging | Pass |

Sterilization:

This lot has been irradiated and dosimetrically released based on the Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization. Sterility Assurance Level: SAL 10⁻⁵

Pyrogens:

Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing". The acceptance level for product is ≤ 0.10 EU/ml or ≤ 4 EU/device.

DNase/RNase:

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10⁻⁷ Kunitz units/uL for DNase and 10⁻⁹ Kunitz units/uL for RNase.

Lot Number Designation:

8 Digit Lot Number: First digit – Last number of year of manufacture, Next 3digits – Julian date, start of manufacturing, Last 3 digits – Batch identification

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

**Quality Assurance Department
Fisher Scientific Company L.L.C.**

This document has been produced electronically and is valid without a signature.

*This certificate has been compiled with data that has been provided by the manufacturer of the product.
Fisher Scientific Channel has not independently verified such data.*